

## Protect and Refresh — Your Breathing Air from Standard Compressors

Air Filtration Process - AirGuard uses state of the art, threestage filtration train configuration to purify the compressed air.

Stage One - Removes 95% of solid contaminants down to 5 microns level


Stage Two - Removes fine oil mists down to 0.01 microns

Stage Three - Removes unpleasant hydrocarbon odors by carbon absorption.

These filters are easily accessible and replaceable by a twist of the outside enclosure of the filter train.

AirGuard uses an intuitive graphical user interface with icons to access instrument functions. A green "Check" icon is displayed if everything is safe and operating as expected. That can change to a warning yellow "Check" if a function needs proactive attention or a red "X"or fault if immediate attention is required.





AirGuard measures the following components in the compressed air stream.

- 1. Carbon Monoxide
- 2. Differential Dew Point
- 3. CFM of the airflow
- 4. Compressor air flow to the sensors

AirGuard uses an electrochemical sensor for the detection of Carbon Monoxide with a range of 0-50 ppm. Additionally, Air-Guard has a "Differential Dew Point" sensor to report the level of humidity/water in the air stream in degrees (°F). The reading indicates the difference in temperature required to create a condensing atmosphere. This feature offers the operator an assessment of the humidity in the compressed air stream and provides the operator with an alarm when there is the possibility of water condensing in the respiratory airline.

AirGuard has an integral CFM measurement of airflow that will provide an alarm should the air flow to the respirator drop below the threshold set-point. This measurement offers the operator an additional level of safety should the air flow decrease unexpectedly due to a compressor or respiratory line failure.

From the main display selecting the check function,
AirGuard will report "System Status" indicating any faults,
CO and differential dew point status, last calibration date,
respiratory flow, and sensor flow measurement.







Record Keeping - In all safety air monitoring applications, keeping records is critical for OSHA compliance. With AirGuard this occurs automatically every time it is powered up. AirGuard collects a comprehensive data log recording both "Carbon Monoxide" and "Differential Dew Point" values, calibration events and diagnostic data on performance. This data is downloadable to a PC from a removable USB drive.



By selecting the "Gear Icon" the Main Menu Settings are displayed allowing access to AirGuard program settings.





## **SPECIFICATIONS**

**Display:** Backlit, 7" diagonal color display

Alarms: User programmable Air Flow Alarm Pt: < 5 CFM ±1 CFM

**Sensor Types:** Carbon Monoxide (0 to 50 ppm)

Dew Point (-20°F to ambient)

**Sensor Accuracy:** Carbon Monoxide: ±5% @ 10 ppm

Dew Point: ±2°F @ 50°F

**Sensor Response Time:** <60 seconds

**Keypad:** 3-button touch, weatherproof w/tactile feedback

**Battery Life:** >16 hours before recharge

**Data Logging:** 2 points per minute **Data Storage Medium:** USB memory stick

Fittings: Hansen or Schrader Connections

Inlet Pressure: Maximum 150 psi
Total Air Capacity: Maximum 15-100 CFM

**Outlet Connection:** Supplied with one to eight couplings

Inlet Port: 1/2 inch quick disconnect Filter Train: 3-stages of filtration

(5 micron, 0.01 micron, carbon adsorber)

Warranty: 2 years on sensors and monitor

(excluding replacement filters)

**Agency Approval:** UL 61010 U.S. and Canada (Pending)

## **OUR MISSION**

ENMET has been in the forefront of creative gas detection solutions since its inception in 1970. Founded by Dr. Verne R. Brown as Environmental Metrology Corporation and abbreviated later as ENMET, our focus has been to develop instruments for industrial health and safety monitoring. We are committed to providing gas detection solutions and services that conform to customer requirements. ENMET operates under the quality registrations for both industrial (ISO-9001:2015) and Aerospace (AS 9100D:2016) certificates.

## **DISTRIBUTED BY:**



680 Fairfield Ct., Ann Arbor, MI 48108 800-521-2978 734-761-1270 www.enmet.com